

PatchTool Compare-1 (#300) (ICC: PA271W OZ101758UA Mult...)

Patch view | Tabular data | Stats

Re-map ▶ | Export... | More info

Ill.: D50 | Obs.: 2 deg. | Color difference: >1.0 | >3.0 | Delta: E* | Formula: CIEDE2000

Rows: 25 | In file: 25 | Lightbox: OFF

Info

File : 1: Mesures P2P51... Density M0.btx
Name : A13
ID : 13

File : 2: Mesures P2P25... Juin 2019).btx
Name : A13
ID : 13

| | | |
|----|--------|--------|
| L* | 55.07 | 55.65 |
| a* | -37.31 | -37.05 |
| b* | -46.75 | -46.98 |

D50, 2 deg. | D50, 2 deg.

FD-7 | **528**

| | | |
|----|------|------|
| C: | N.A. | N.A. |
| M: | N.A. | N.A. |
| Y: | N.A. | N.A. |
| K: | N.A. | N.A. |

Density: Status T

ΔE* CIEDE2000: **0.57**

| | |
|-------|-------|
| ΔL*/w | 0.55 |
| Δa' | 0.26 |
| Δb' | -0.23 |

Info

File : 1: Mesures P2P51... Density M0.btx
Name : B13
ID : 38

File : 2: Mesures P2P25... Juin 2019).btx
Name : B13
ID : 38

| | | |
|----|-------|-------|
| L* | 48.66 | 49.48 |
| a* | 74.25 | 74.10 |
| b* | -2.44 | -2.32 |

D50, 2 deg. | D50, 2 deg.

FD-7 | **528**

| | | |
|----|------|------|
| C: | N.A. | N.A. |
| M: | N.A. | N.A. |
| Y: | N.A. | N.A. |
| K: | N.A. | N.A. |

Density: Status T

ΔE* CIEDE2000: **0.82**

| | |
|-------|-------|
| ΔL*/w | 0.82 |
| Δa' | -0.15 |
| Δb' | 0.12 |

Info

File : 1: Mesures P2P51... Density M0.btx
Name : C13
ID : 63

File : 2: Mesures P2P25... Juin 2019).btx
Name : C13
ID : 63

| | | |
|----|-------|-------|
| L* | 86.84 | 87.77 |
| a* | -4.15 | -4.56 |
| b* | 94.57 | 95.44 |

D50, 2 deg. | D50, 2 deg.

FD-7 | **528**

| | | |
|----|------|------|
| C: | N.A. | N.A. |
| M: | N.A. | N.A. |
| Y: | N.A. | N.A. |
| K: | N.A. | N.A. |

Density: Status T

ΔE* CIEDE2000: **0.65**

| | |
|-------|-------|
| ΔL*/w | 0.60 |
| Δa' | -0.41 |
| Δb' | 0.87 |

Info

File : 1: Mesures P2P51... Density M0.btx
Name : D25
ID : 100

File : 2: Mesures P2P25... Juin 2019).btx
Name : D25
ID : 100

| | | |
|----|-------|-------|
| L* | 13.16 | 13.77 |
| a* | 0.51 | 0.53 |
| b* | 2.35 | 2.24 |

D50, 2 deg. | D50, 2 deg.

FD-7 | **528**

| | | |
|----|------|------|
| C: | N.A. | N.A. |
| M: | N.A. | N.A. |
| Y: | N.A. | N.A. |
| K: | N.A. | N.A. |

Density: Status T

ΔE* CIEDE2000: **0.41**

| | |
|-------|-------|
| ΔL*/w | 0.40 |
| Δa' | 0.03 |
| Δb' | -0.11 |

Info

File : 1: Mesures P2P51... Density M0.btx
Name : B1
ID : 26

File : 2: Mesures P2P25... Juin 2019).btx
Name : B1
ID : 26

| | | |
|----|-------|-------|
| L* | 91.05 | 92.19 |
| a* | 0.52 | 0.36 |
| b* | 1.16 | 1.43 |

D50, 2 deg. | D50, 2 deg.

FD-7 | **528**

| | | |
|----|------|------|
| C: | N.A. | N.A. |
| M: | N.A. | N.A. |
| Y: | N.A. | N.A. |
| K: | N.A. | N.A. |

Density: Status T

ΔE* CIEDE2000: **0.78**

| | |
|-------|-------|
| ΔL*/w | 0.70 |
| Δa' | -0.24 |
| Δb' | 0.26 |

Info

File : 1: Mesures P2P51... Density M0.btx
Name : A25
ID : 25

File : 2: Mesures P2P25... Juin 2019).btx
Name : A25
ID : 25

| | | |
|----|-------|-------|
| L* | 47.59 | 48.15 |
| a* | 68.63 | 68.56 |
| b* | 50.32 | 50.67 |

D50, 2 deg. | D50, 2 deg.

FD-7 | **528**

| | | |
|----|------|------|
| C: | N.A. | N.A. |
| M: | N.A. | N.A. |
| Y: | N.A. | N.A. |
| K: | N.A. | N.A. |

Density: Status T

ΔE* CIEDE2000: **0.57**

| | |
|-------|-------|
| ΔL*/w | 0.55 |
| Δa' | -0.07 |
| Δb' | 0.35 |

Info

File : 1: Mesures P2P51... Density M0.btx
Name : B25
ID : 50

File : 2: Mesures P2P25... Juin 2019).btx
Name : B25
ID : 50

| | | |
|----|--------|--------|
| L* | 48.37 | 48.90 |
| a* | -63.58 | -62.93 |
| b* | 32.23 | 32.30 |

D50, 2 deg. | D50, 2 deg.

FD-7 | **528**

| | | |
|----|------|------|
| C: | N.A. | N.A. |
| M: | N.A. | N.A. |
| Y: | N.A. | N.A. |
| K: | N.A. | N.A. |

Density: Status T

ΔE* CIEDE2000: **0.56**

| | |
|-------|------|
| ΔL*/w | 0.52 |
| Δa' | 0.65 |
| Δb' | 0.08 |

Info

File : 1: Mesures P2P51... Density M0.btx
Name : C25
ID : 75

File : 2: Mesures P2P25... Juin 2019).btx
Name : C25
ID : 75

| | | |
|----|--------|--------|
| L* | 23.45 | 23.80 |
| a* | 21.93 | 22.06 |
| b* | -45.69 | -45.54 |

D50, 2 deg. | D50, 2 deg.

FD-7 | **528**

| | | |
|----|------|------|
| C: | N.A. | N.A. |
| M: | N.A. | N.A. |
| Y: | N.A. | N.A. |
| K: | N.A. | N.A. |

Density: Status T

ΔE* CIEDE2000: **0.30**

| | |
|-------|------|
| ΔL*/w | 0.26 |
| Δa' | 0.13 |
| Δb' | 0.15 |

PatchTool Compare-1 (#300) (ICC: PA271...)

Patch view | Tabular data | Stats

More bins! | Less bins! | bin size: 0.100 | CRF

| | | |
|--------------|------|--|
| Average | | |
| All samples: | 0.75 | |
| Best 90%: | 0.70 | |
| Worst 10%: | 1.14 | |

| | | |
|--------------|------|--|
| Sigma | | |
| All samples: | 0.22 | |
| Best 90%: | 0.15 | |
| Worst 10%: | 0.34 | |

Delta: E* | Formula: CIEDE2000

Absolute | L*C*H* weighted | Separate Neg / Pos stats

| | | |
|--------------|--------|--|
| Max error | | |
| 10th perc: | ± 0.53 | |
| Median: | ± 0.74 | |
| 90th perc: | ± 0.98 | |
| 95th perc: | ± 1.04 | |
| All samples: | ± 2.87 | |

Stats report | No. samples: 0 + 300 = 300

Minolta FD-7 compared to X-Rite 528
P2P51 IP 18pts Everest SBS / June 11 2019 / Roger Breton
www.graxx.ca